Thermometer Hipot 32 pin Connector

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT		03/17/1999

Test Date: 06/17/2005

Test Time: 12:14:48 AM

User Name: [illegible]

Device Tested

Model Number: BOX A

Serial Number: CERNX TEMP. SENSORS 32 PIN

Setup Filename: Leads In Air & He.

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	9	3	Pass	Pass	60.0s	0.30KV	0.0microamps