Leak Test: Pressurized Internal Circuits with Vacuum Vessel Evacuated

Meyer Tool & Mfg., Inc.
4601 SOUTHWEST HWY., OAK LAWN, ILLINOIS 60453
708 / 425-9080
FAX 708 / 425-2612

PNEUMATIC PRESSURE TEST REPORT

MTM JOB NO.:	00918
CUSTOMER:	LBNL
MTM DRAWING/ASSEMBLY/ITEM NO.:	24350
COMPONENT SERIAL NUMBER :	Box "B"
COMPONENT DESCRIPTION:	FEED BOX
PRESSURE GAUGE(S) TAG NO.:	PG-26
PRESSURE GAUGE(S) RANGE:	0/60 psi
TEST PRESSURE:	28.5 psig
GAS USED:	HELIUM
PROCEDURE USED:	☑ MTM-MP-1105 ☐ OTHER _____
RESULTS/COMMENTS:	VAC. SHELL UNDER VACUUM, PRESSURIZED MOX2, JC1 & JC2 SEPARTLY. NO LOSS OF PRESSURE – NO LEAKS DETECTED ON HMOD.
OPERATOR:	Wayne E. [illegible] DATE: 7-13-05 ☑ ACCEPT ☐ REJECT

TEST REPORT FORM TR-1105
REV. O