Leak Test: Pressurized Internal Circuits with Vacuum Vessel Evacuated

Meyer Tool & Mfg., Inc.
4601 SOUTHWEST HWY., OAK LAWN, ILLINOIS 60453
708 / 425-9080
FAX 708 / 425-2612

PNEUMATIC PRESSURE TEST REPORT

MTM JOB NO.:	00918
CUSTOMER:	LBNL
MTM DRAWING/ASSEMBLY/ITEM NO.:	24C352
COMPONENT SERIAL NUMBER :	Box "C"
COMPONENT DESCRIPTION:	FEED BOX
PRESSURE GAUGE(S) TAG NO.:	PG-26
PRESSURE GAUGE(S) RANGE:	0 / 60 psi
TEST PRESSURE:	28.8 psig
GAS USED:	HELIUM
PROCEDURE USED:	☑ MTM-MP-1105 ☐ OTHER VACUUM SHELL UNDER VACUUM, PRESSURIZED: JC2 CIRCUIT, JC1 CIRCUIT, MBX2 CIRCUIT MQX2 CIRCUIT.
RESULTS/COMMENTS: JC1 CIRCUIT CONTAINS: XB, E1, LD2, LD1 EX JC2 CIRCUIT CONTAINS: OH, CY2, CC3, E2, LD3, CC1 (JACKET) (LHE TANK) (MQX1) (MQX2)	NO LOSS OF PRESSURE, NO INDICATION OF LEAK
OPERATOR:	Wayne Eghan DATE: 9-13-05 ☑ ACCEPT ☐ REJECT