Thermometer Hipot 32 pin Connector

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 10/26/2005

Test Time: 06:14:17 PM

User Name: Glen Krecky

Device Tested

Model Number: box C

Serial Number: cernox therms. 32 pin connector

Setup Filename: Leads In Air & He.

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	28	1	Pass	Pass	60.0s	0.30KV	0.0microamps