Thermometer Hipot 32 pin Connector

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 10/22/2004

Test Time: 01:46:23 AM

User Name: SRO 10/21/04

Device Tested

Model Number: RTD Cryo Pipes 32 pin IN AIR

Serial Number: Box D

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	32	1	Pass	Pass	60.0s	0.20KV	0.0microamps
~~2~~	~~AC Withstand~~	~~32~~	~~2~~	~~Fail~~	~~Abort~~	~~0.5s~~	~~1.20KV~~	~~0.000mA~~

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	01/31/2004

Test Date: 12/07/2004

Test Time: 10:34:34 PM

User Name: D. [illegible]

Device Tested

Model Number: Temp. Sensor Connector - Cryo Pipe Rds 32 Pin

Serial Number: Box D (Air - Helium not possible)

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	30	1	Pass	Pass	60.0s	0.20KV	0.0microamps