[illegible]	[illegible]
[illegible]	24 C394
COMPONENT SERIAL NUMBER:	BOX "E"
COMPONENT DESCRIPTION:	[illegible] LINE
PRESSURE GAUGE(S) [illegible]	[illegible]
PRESSURE GAUGE(S) [illegible]	[illegible] PSI
TEST PRESSURE:	364 PSI
GAS USED:	[illegible]
PROCEDURE USED:	☑ MTM-MP-1105
	☐ OTHER NO LOSS OF PRESSURE
RESULTS/COMMENTS:	WITH IN 10 MINUTES
OPERATOR:	[signature] DATE: 2/23/05
	☒ ACCEPT
	☐ REJECT

TEST REPORT FORM TR-1105
REV. O

MTM JOB NO.:	00918
CUSTOMER:	LBNL
MTM DRAWING/ASSEMBLY/ITEM NO.:	24 C 394
COMPONENT SERIAL NUMBER:	BOX "E"
COMPONENT DESCRIPTION:	MQX2 LINE
MSLD TAG NO.:	3094
CALIBRATED LEAK TAG NO.:	CL-1
SENSITIVITY ([X] MSLD [ ] SYSTEM):	INITIAL 9.333X10 / FINAL 9.333X10
TEST PERFORMED ON SCALE:	100
INITIAL READING:	90/100
FINAL READING:	90/100
LEAK RATE (ATM CC/SEC)	
PROCEDURE USED:	[X] MTM-MP-1110 [ ] OTHER NO LEAKS
RESULTS/COMMENTS: (NOTE: Record hood test time vs. MSLD reading in this section.)	DETECTED
OPERATOR:	Glen D. Koreck DATE: 2/23/05
	[X] ACCEPT [ ] REJECT

TEST REPORT FORM TR-1110
REV. O