Leak Test: Pressurized Internal Circuits with Vacuum Vessel Evacuated

Meyer Tool & Mfg., Inc.
4601 SOUTHWEST HWY., OAK LAWN, ILLINOIS 60453
708 / 425-9080
FAX 708 / 425-2612

PNEUMATIC PRESSURE TEST REPORT

MTM JOB NO.:	00918
CUSTOMER:	LBNL
MTM DRAWING/ASSEMBLY/ITEM NO.:	290395
COMPONENT SERIAL NUMBER :	BOX "F"
COMPONENT DESCRIPTION:	FEED BOX
PRESSURE GAUGE(S) TAG NO.:	PG 26
PRESSURE GAUGE(S) RANGE:	0-60 PSI
TEST PRESSURE:	28.5 PSIG
GAS USED:	He.
PROCEDURE USED:	☒ MTM-MP-1105 ☐ OTHER WITH VACUUM SHELL UNDER VACUUM. 1 PRESSURIZING
RESULTS/COMMENTS: NOTE: JC1 CIRCUIT CONTAINS XB, LD2 LD1 V- JACKED, E1 MRX1 JC2 CIRCUIT CONTAINS DH, CC3, CC1, CC2, KD2 EX (E2), V	MRX2, JC1, AND JC2 CIRCUITS SEPARTLTY, DETECTED NO LOSS OF PRESS, NOR DID I DETECT ANY RISE OH MY MSLD
OPERATOR:	Glenn Moody DATE: 5/31/05
	☒ ACCEPT ☐ REJECT