MBX2 Hipot in Helium

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:33:23 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: 19 pin connector 1-8

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	0.20KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:24:27 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: M5, M6, M7

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	1.40KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:22:06 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: N6, N7

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	1.40KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:19:56 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: N4, N5

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	1.40KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:14:31 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: N2, N3

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	0.65KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:11:57 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: M4, N1

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	0.65KV	0.0microamps

Test Instruments Used

Test Model #	Test Serial #	Calibration Due Date
7550DT	9111100	02/01/2004

Test Date: 09/30/2005

Test Time: 07:08:04 PM

User Name:

Device Tested

Model Number: Feed Box H MBX2 in He

Serial Number: M1, M2, M3

Setup Filename:

Results

	Test Type	Memory	Step	Pass/Fail	Test Result	Timer	Parameter 1	Parameter 2
1	DC Withstand	21	1	Pass	Pass	60.0s	1.40KV	0.0microamps